

EC DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'CE' MARK



We, the manufacturer: BIOQUELL UK Limited.
52 Royce Close,
West Portway
Andover, England
SP10 3TS
Tel: +44 (0) 1264 - 835835
Fax: +44 (0) 1264 - 835836

declare under our sole responsibility that the appliance:

BQ-50

Serial Number as stated on the unit's rating plate

complies with the Essential Requirements of the Machinery Directive 2006/42/EC, Radio Equipment Directive 2014/53/EU, and Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU by means of the following harmonised standards:

Machinery Directive 2006/42/EC: by means of EN 12100:2010

Radio Equipment Directive 2014/53/EU by means of:

Article 3.1a	RF Exposure: EN 62311
Article 3.1a	EN 61010-1: 2010
Article 3.1b	EN 61326-2-1:2013 using ETSI EN 301 489-3 V2.1.1, & ETSI EN 301 489-17 V3.1.1 (ETSI EN 301 489-1 V2.1.1)
Article 3.2	EN 300 328 v 2.1.1 EN 300 330 v 2.1.1

Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU by means of EN 50581:2012

A copy of the Technical File for this equipment is available from:

O.R.Cumberlege

At Bioquell UK Limited, 52 Royce Close, West Portway, Andover, England

Authorised signatory of manufacturer:

O.R.Cumberlege, Director

10th January 2019